



851 Transport Drive Phone: 219 465 2700  
Valparaiso, IN 46383-8432 Web: www.gasleaksensors.com

— Innovative Detection Solutions —

**MADE IN THE USA**  
WITH GLOBALLY SOURCED COMPONENTS

Distributed by:

## Quick-Start Instructions



This is not a substitute for the instruction manual. It is for reference only.

Read and understand the instruction manual before use.

# Trak-It® IIIa GLT

## QUICK-START INSTRUCTIONS

1. Turn on TRAK-IT IIIa GLT Device and Tablet (for the tablet press and hold the top left button).
2. Wait until TRAK-IT IIIa GLT displays 'WAITING FOR TRAINER' on screen.
3. Note: If TRAK-IT IIIa GLT Device has passed 'WAITING FOR TRAINER' screen (Either you have skipped or it has already been connected with application), It will take you directly to the Instrument screen.
4. Launch the GLT application (if it does not auto launch).
5. Tap the "Connect" on the device from the list
6. After the connection has been made, tap either of the two options
7. Training Mode: It will put device in simulation mode and allows trainer to control device (sensors are inactive).
8. Monitoring Mode: It will put device in Normal mode and allows trainer to monitor remotely (sensors are active).
9. The TRAK-IT IIIa GLT will continue through warm-up and arrive at the work display.
10. The tablet will display the TRAK-IT IIIa GLT screen with adjustable slide bars if in "training mode" or just the screen if in "monitor mode".
11. The training modules will be displayed simultaneously on the right hand side or bottom side of the screen depending on the orientation of the tablet.
12. Note: If you need to switch the mode, you need to 'Disconnect' and Restart the 'TRAK-IT IIIa GLT'.
13. Note: If any configurations or mode changes on 'TRAK-IT IIIa GLT', please power cycle the TRAK-IT IIIa GLT and connect to the application again.

